# IMT. Analytics

## **EU Declaration of Conformity**

Herewith, the manufacturer

### IMT Analytics AG

Gewerbestrasse 8 9470 Buchs SG Switzerland

declares on sole responsibility that the production and the distribution of the following products:

Description	Type
Flow-Pressure Measurement Device	CITREX H3
Flow-Pressure Measurement Device	CITREX H4

were made according to the quality rule EN ISO 13485:2016 (quality management systems) and applicable Directives:

### 2015/863/EU

Directive (EU) 2015/863 (RoHS-3) amending Annex II to Directive 2011/65/EU (RoHS-2) as regards restriction of Hazardous Substances in Electrical and Electronic Equipment. To achieve this, we certify based on available information conducted to us, the product identified above has never contained or has never been contaminated during fabrication process by the following homogeneous materials above the allowable quantity limits according to the Directive: Lead (0,1 %), Mercury (0,1 %), Cadmium (0,01 %), Hexavalent chromium (0,1 %), Polybrominated biphenyls (PBB) (0,1 %), Polybrominated diphenyl ethers (PBDE) (0,1 %), Bis(2-ethylhexyl) phthalate (DEHP) (0,1 %), Butyl benzyl phthalate (BBP) (0,1 %), Dibutyl phthalate (DBP) (0,1 %)

### 2014/35/EU

Directive 2014/35/EU of the European Parliament and of the Council of 26 February 2014 on the harmonization of the laws of Member States relating to the making available on the market of electrical equipment designed for use within certain voltage limits.

### 2014/30/EU

Directive 2014/30/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to electromagnetic compatibility.

EN 61326-1: 2013	Electrical equipment for measurement, control and laboratory use - EMC requirements
EN 61010-1: 2010, EN 61010-1:2010/A1:2019	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements
EN 61010-1:20107A1:2019/AC:2019-04	

The conformity of the products is validated according to abovementioned directives and by the application of the sign



Signed for and on behalf of: IMT Analytics AG

Place and date of issue: Buchs SG, 22. August 2022

Name, function: Hans-Jörg Reinhold, Senior

Regulatory Affairs Manager

Signature:

21! Rein

# IMT. Analytics

## **EU Declaration of Conformity**

Herewith, the manufacturer

### IMT Analytics AG

Gewerbestrasse 8 9470 Buchs SG Switzerland

declares on sole responsibility that the production and the distribution of the following product:

Description	Type
Flow-Pressure Measurement Device	CITREX H5

were made according to the quality rule EN ISO 13485:2016 (quality management systems) and applicable Directives:

### 2015/863/EU

Directive (EU) 2015/863 (RoHS-3) amending Annex II to Directive 2011/65/EU (RoHS-2) as regards restriction of Hazardous Substances in Electrical and Electronic Equipment. To achieve this, we certify based on available information conducted to us, the product identified above has never contained or has never been contaminated during fabrication process by the following homogeneous materials above the allowable quantity limits according to the Directive: Lead (0,1 %), Mercury (0,1 %), Cadmium (0,01 %), Hexavalent chromium (0,1 %), Polybrominated biphenyls (PBB) (0,1 %), Polybrominated diphenyl ethers (PBDE) (0,1 %), Bis(2-ethylhexyl) phthalate (DEHP) (0,1 %), Butyl benzyl phthalate (BBP) (0,1 %), Dibutyl phthalate (DBP) (0,1 %), Diisobutyl phthalate (DIBP) (0,1 %)

#### 2014/35/EU

Directive 2014/35/EU of the European Parliament and of the Council of 26 February 2014 on the harmonization of the laws of Member States relating to the making available on the market of electrical equipment designed for use within certain voltage limits.

#### 2014/30/EU

Directive 2014/30/EU of the European Parliament and of the Council of 26 February 2014 on the harmonisation of the laws of the Member States relating to electromagnetic compatibility.

EN 61326-1: 2013	Electrical equipment for measurement, control and laboratory use - EMC requirements
EN 61010-1: 2010 EN 61010-1:2010/A1:2019	Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements
EN 61010-1:20107A1:2019/AC:2019-04	

### 2014/53/EU

Directive 2014/53/EU of the European Parliament and of the Council of 16 April 2014 on the harmonization of the laws of the Member States relating to the making available on the market of radio equipment.

EN 301 489-1 v2.1.1 and EN 301 489-17 v3.1.1	EMC for broadband data transmission systems (Article 3.1(b))
EN 300 328 v2.2.2	SPECTRUM for wideband transmission systems in 2.4 GHz ISM band (Article 3.2)

The conformity of the product is validated according to abovementioned Radio Equipment Directive 2014/53/EU, by making reference to Directives 2014/30/EU and 2014/35/EU and by the application of the sign



Signed for and on behalf of: IMT Analytics AG

Place and date of issue: Buchs SG, 22. August 2022

Name, function: Hans-Jörg Reinhold, Senior Regulatory Affairs Manager

Signature:

MI-Oleun